



**SHRI GURU GOBIND SINGHJI INSTITUTE OF
ENGINEERING&TECHNOLOGY, NANDED-06**

File No. SGGsIE&T/Store-2679/Electro Discharge Machine & Coating Thickness Measurement/Prod/2017-18

Corrigendum

Refer our e-tender id. **2017_DTEDM_253084_1** Which was e-tender notice published in **Daily Sakal, Nanded** on dated **08.11.2017**, The tender is available on portal <https://mahaetenders.gov.in> as well as Institute web site <http://sggs.ac.in>.

The name of the machine coating thickness measurement machine we have mentioned in tender please read the changes as per given below.

Sr. No.	Specifications by SGGsIE&T	Specification as in tender notice	Specifications required
01	Range : Thickness range	0.1 μ m to 300 μ m	0.1 μ m to 50 μ m
02	Measurement : Accurate measurement by different equipments Ex. Balls etc	dia 10 to 100 mm	Measurement: Accurate measurement by different diameters of balls dia 10 to 50 mm
03	Speed	100 to 10000 rpm	Motor speed 300 to 3000 rpm
04	Time	1 to 30000 sec	Time 1 to 10000 sec
05	Thickness : Thickness of coating	0.1 to 300 μ m	Thickness Range 0.1 to 50 μ m

**Sd/-
DIRECTOR**